


FlowAccessV3

Date:2014/9/17 15:04

CN
I Order ISO 8466



$a = -164.48634812286800$ $b = 2172.65614334471000$ $RSD = 914.61481989487900$

$r = 0.99993649984534$ $R2 = 0.99987300372295$

Run Name : SKALAR20140917A0

User Name : Administrator

操作员名字 : Administrator

FlowAccessV3 Results Report

Run Name : SKALAR20140917A0

DateTime :2014/9/17 12:02:00

用户名 : Administrator

操作员名字 : Administrator

	NeedleNumber	ResultID	Position	SampleType	SampleIdentity	CN- Results	CN-CorrectedHeight
1	1	0	IW	IW	InitialWash	0.08	0
2	1	1	A1	T	Tracer	104.12	226054
3	1	2	WT	W	Wash	0.08	0
4	1	3	WT	S1	0.00	0.08	14
5	1	4	A2	S2	10	10.28	22167
6	1	5	A3	S3	20	19.79	42842
7	1	6	A4	S4	40	39.88	86481
8	1	7	A5	S5	60	60.14	130489
9	1	8	A6	S6	80	79.25	172014
10	1	9	A7	S7	100	100.58	218365
11	1	10	A8	D	Drift 40.00	42.04	91177
12	1	11	WT	W	Wash	0.08	0